

BUS/SCS nurse call unit

signalling and acknowledgement units for rooms

Antimicrobial⁽¹⁾



0 766 06



0 766 07



0 766 08



0 782 00L



Selection chart for boxes, supports and plates **p. 1198-1199**

BUS/SCS system with star wiring around the door unit:

- call centralisation
- acknowledgement
- communication with nursing control unit and all elements of the system via BUS/SCS
- signalling forwarded to all indicators (e.g.: corridor overhead light unit, door unit, hand-held remote control unit)

Pack	Cat.Nos	Antimicrobial ⁽¹⁾ door units for rooms
1	0 766 06	<p>For installation at the entrance to the room, can be seen from the bed</p> <p>Connect directly to the BUS/SCS cable</p> <p>Cat.Nos 0 492 32/33</p> <p>With indicator and pushbutton for calls, nurse presence and 2nd nurse presence</p> <p>Comply with standards VDE 0834-1 and 0834-2</p> <p>Fault display</p> <p>4-module wiring mechanisms supplied complete</p> <p>Flush mounting in 50 mm deep box or surface mounting with frame Cat.No 0 766 14 (p. 1211)</p> <p>Can be acknowledged by a pushbutton with indicator Cat.No. 0 782 04</p> <p>Door unit with indicators</p> <p>Door unit with illuminated call indicators</p> <p>Nurse presence and acknowledgement via pushbutton. Audible and illuminated signal of a call coming from another room. Suitable for shared bathrooms and for secure wandering</p> <p>Door unit with display</p> <p>Door unit with alphanumeric display. Nurse presence and acknowledgement via pushbutton. Call display from another room. Used in conjunction with the interphone unit Cat.No 0 766 09 for communicating with rooms issuing calls and with a nurse present. User is able to acknowledge at the end of the communication</p> <p>Bed extension</p> <p>To be associated with Cat.No 0 766 06 or 0 766 07</p> <p>Allows the connection of 4 components to 1 call unit DIN, 6 modules. Programming via virtual configuration</p>
1	0 766 07	<p>Antimicrobial⁽¹⁾ interphone unit for room and nurses' station</p> <p>Interphone door unit</p> <p>For installation:</p> <ul style="list-style-type: none"> - at the entrance to the room connected to the door unit with display Cat.No 0 766 07 - in the nurses' room linked to the control unit Cat.No 0 766 11 (p. 1211) <p>Allows the user to:</p> <ul style="list-style-type: none"> - communicate from one room to another room issuing a call or with a nurse present - broadcast a general call <p>Must be connected to a microphone Cat.No 0 782 00 for communication between the patient and the nursing staff</p> <p>4-module wiring mechanism supplied complete</p> <p>Flush mounting for Mosaic in 50 mm deep box or surface mounting with frame Cat.No 0 766 14 (p. 1211)</p> <p>Room microphone</p> <p>Remote microphone for connection to the interphone unit Cat.No 0 766 08</p> <p>For installation next to the bed. For communication between the patient and the nursing staff. 2-module mechanism. Supplied pre-assembled with white inishing plate and support frame</p>
1	0 782 00L	

BUS/SCS nurse call unit

corridor signalling units



0 766 70



0 766 04



0 766 71L



Selection chart for boxes, supports and plates **p. 1198-1199**

Ensures that information is transferred into the corridor

Pack	Cat.Nos	Corridor signalling devices
1	0 766 71L	<p>For installation in the corridor above the doors</p> <p>Supplied with support and white finishing plate</p> <p>Flush-mounting or surface-mounting.</p> <p>Call only overdoor light unit for corridors</p> <p>LED triangular overdoor light unit for call signalling</p> <p>1 red indicator</p> <p>2 modules</p> <p>Call and nurse presence overdoor light units for corridors</p> <p>LED triangular overdoor light units for call signalling, bathroom calls and nurse presence</p> <p>Monobloc indicator in red, white, green</p> <p>2 modules</p> <p>Recommended for compliance with standards VDE 0834-1 and 0834-2</p> <p>Corridor display unit</p> <p>For installation in the corridor</p> <p>10-character alphanumeric display</p> <p>Indication of calls and nurse presence by level of priority</p> <p>For fixing in false ceiling or on the wall</p> <p>Supplied with connection interface to the BUS/SCS</p> <p>Double sided display unit</p>
1	0 766 70	
1	0 766 04	
1	0 766 05	
1	0 766 03	<p>Fixing accessory for corridor display unit</p> <p>For fixing a corridor display unit projecting</p>
1	0 766 78	<p>Transfer of information</p> <p>Interface to transfer information to an indicator (Cat.No 0 766 71L) or to an electronic bell (Cat.No 0 766 42)</p>
1	0 766 42	<p>Electronic bell</p>

1: Contains a silver compound to prevent the growth of bacteria on the surface of the plate